

OF-3-162

Use and Care of Your Hearing Instruments



ACCESSORIES



Storage case



Cleaning tool

Your hearing instruments are shipped with a variety of accessories, a sampling pictured above, including the following:

- Storage case
- 1 box of batteries per instrument
- Storing/Drying jar
- Program selection wand
- Cleaning tools

If you have any questions about the use of these accessories please contact your hearing care professional.

TUBING REPLACEMENT



Note: It is important **NOT** to pull the tubing itself, as it can become stretched and distorted requiring replacement.




REMOVAL: Grasp the connector firmly. Do not pull the tubing itself, as it can become stretched and distorted. Twist the connector 90 degrees (as shown.) Pull straight out.



ATTACHMENT: To attach, align the acoustic tube plug with the opening and push firmly until it snaps into place. Note: You will feel a firm de-dent as it locks into place.


PARTS OF YOUR OF-3-162



Inside Battery Door
Blue = Left Side



Inside Battery Door
Red = Right Side



OF-3-162 ANSI SPECIFICATIONS	
MAX SSPL 90	130 dB +/- 4
HFA SSPL 90	128dB +/- 4
MAX FOG	50dB +/- 5
HFA FOG	45dB +/- 5
THD 500Hz	≤1% MAX
800 Hz	≤1% MAX
FREQ RANGE F1	1.6kHz ≤2% MAX
FREQ RANGE F2	200Hz
F2	8000Hz
REF TEST GAIN	43dB +/- 3
EN	≤23dB MAX
BAT DRAIN	1.25mA MAX

BATTERY INSERTION



Your hearing aids are designed to use size 13 batteries. A zinc-air type battery is recommended, but any type or brand may be used.

To insert, remove the adhesive label and wait thirty seconds for air to activate the battery. Any moisture on the battery surface should be wiped off prior to insertion. **With the positive side up**, insert the battery into the battery door. It may take several minutes for the battery to work to its full effect.



To remove the battery, open the battery door and lift the battery out of the battery door.



Low battery indicator: When your battery voltage is low and in need of replacement you will hear the low battery indicator. You will hear five beeps indicating that it is time to change your battery. The warning beeps will be repeated.




America Hears, Inc.

806 Beaver Street Bristol, PA 19007
800-492-4515

INSERTING HEARING AIDS

Inserting your hearing instruments for the first time can be challenging. Until your hearing aid is fully inserted there can be feedback (whistle). We suggest that you practice without the battery until you have perfected this procedure. With practice, inserting your hearing aids will become as easy as putting on your glasses.



To insert your hearing instrument:

Grasp the hearing aid between your thumb and forefinger.

Place it over your ear as shown.

Gently push the ear piece into the ear canal. Tuck the concha lock into the bowl of your ear.

When positioned correctly the your hearing aid is almost completely invisible in your ear.

Note: It is important to remove the instrument by grasping the instrument itself. Please do NOT remove the instrument by pulling the tubing, as it can become stretched and distorted and may require replacement.

ON/OFF FUNCTION & VOLUME

Your hearing aid is turned on and off by opening and closing the battery compartment or by turning the volume control to the lowest possible setting.



Note: Whenever you turn your hearing aid on it is automatically set to Program One.

Hint: Turning your instrument off every time you remove it will prolong the life of your hearing aid batteries.

Adjust the volume level of your hearing instruments by rotating the volume control up for louder sound and down for softer sounds.

The lower the number, the softer the sound. The higher the number, the louder the sound.

PROGRAM SELECTION

As you change programs, you will hear a series of beeps, which will represent the change, as listed below



When you turn your hearing aid on it is automatically set to program one.

- Program 1 (General Use): One Beep
- Program 2 (Noisier Situations): Two Beeps
- Program 3 (Windy Situations): Three Beeps
- Program 4 (Not Active): Four Beeps

Use the push button or the program wand to change between programs. To use the wand, touch it to your ear where the hearing aid sits (as shown) and move it away. To change to the next memory, again, bring the wand back up to your ear and move it away.

Warning: Do not place the Program Wand on or near computer equipment, software, discs or credit cards. Information could be lost/destroyed as a result.

GETTING CONNECTED

Shiny Side Connect Flex Connector and Cord Here



Your hearing aids use a flex connector to connect the programming cord to the hearing instruments.

- Align red dots on the flex connector and programming cable.
- Connect programming cable to the flex connector.
- Open the battery door slightly to create a space on the hinged side of the battery door.
- Insert the flex connector into the space created. The gold (shiny) side of the flex connector must be facing away from the battery compartment. Do not force this connection.
- Close the battery door.

Refer to the QuickStart Guide and/or the Virtual Office Tutor for programming instructions.

CLEANING TOOL

Daily wipe your instruments, tubing, sound port, and tips/molds clean with a soft, dry cloth, carefully removing any traces of wax and debris. A brush is provided to further assist you in cleaning.

After being disconnected from your hearing instrument, your tips/molds can be washed in lukewarm water with a mild detergent. Reconnect after drying thoroughly.



To clean your thin acoustic tube detach tube from the instrument and then insert the cleaning tool into the plug end of the tube. Push all the way through the tube, removing wax and debris out of the sound tip end.



STORING/DRYING JAR

We recommend that you store your hearing instruments in the drying kit when the aids are not in use.



Open the battery door of your hearing instruments, place them on top of the circular foam piece, and firmly secure lid on the jar. Removing batteries before storing will extend battery life.

There is a small amount of blue beads in the jar that are designed to fade and blend in with the other beads when it is time to reactivate. After reactivation these beads will return to their original blue color. Keeping the lid firmly in place at all times will extend the time between reactivation.

Carefully follow the directions on the jar to reactivate the beads. Allow the beads to cool to room temperature before placing them back into the jar.

Never put your hearing instruments, batteries, glass jar, lid or foam in the oven or microwave. Do not place over an open flame.